



## OIML Certificate



Number R76/2006-A-NL1-21.36 Project number 2522017 Page 1 of 2

Issuing authority NMi Certin B.V.

Person responsible: M.Ph.D. Schmidt

Applicant and Manufacturer

**Rice Lake Weighing Systems** 230 West Coleman Street Rice Lake, WI 54868 United States of America

Identification of the certified type

A Non-automatic weighing instrument

MSI-3460C

Characteristics See next page

This OIML Certificate is issued under scheme A.

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class (III)

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.



#### Issuing Authority

### NMi Certin B.V., OIML Issuing Authority NL1 5 August 2021



**Certification Board** 

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

verified in the blue ribbon on top of the electronic version of this certificate.





NMi Certin B.V. Thiissewea 11 2629 JA Delft The Netherlands T+31 88 6362332 certin@nmi.nl www.nmi.nl

This document is digitally signed and sealed. The digital signature can be



The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org



# **OIML** Certificate

**OIML Member State**The Netherlands

Number R76/2006-A-NL1-21.36 Project number 2522017 Page 2 of 2

The conformity was established by the results of tests and examinations provided in the associated OIML Test Report(s):

- No. NMi-2522017-01 dated 5 August 2021 that includes 26 pages;
- No. NMi-2522017-02 dated 5 August 2021 that includes 24 pages;
- No. NMi-2522017-03 dated 5 August 2021 that includes 11 pages;
- No. NMi-2522017-04 dated 5 August 2021 that includes 23 pages;
- No. NMi-2522017-05 dated 5 August 2021 that includes 24 pages.

#### **Characteristics of the non-automatic weighing instrument:**

Accuracy class	
Maximum capacity	250 ≤ Max ≤ 5000
Verification scale interval	e ≥ 0,1 kg
Weighing range(s)	Single-interval
Maximum number of scale intervals (one weighing range)	n ≤ 2500
Temperature range	-10 °C / +40 °C
Power supply voltage	6 V DC by rechargable
Software identification	Version number : 05-xx (x = 0 99)