



NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance

for Weighing and Measuring Devices

For:

On-Board Weighing System
Lift Truck Scale, Load Cell Electronic, Static and In-Motion
Model: CLS
 n_{max} : See table below
 e_{min} : See table below
Capacity: 5 000 lb
Accuracy Class: III

Submitted By:

Rice Lake Weighing Systems
230 West Coleman Street
Rice Lake, WI 54868
Tel: 715-234-9171
Fax: 715-234-6967
Contact: Derrick Bender
Email: dbender@ricelake.com
Web site: www.ricelake.com

Standard Features and Options**Operational Parameters for Static Mode Operation**

| Maximum Fork Length (inches) | Capacity (lb) | e_{min} (lb) | n_{max} |
|------------------------------|---------------|----------------|-----------|
| 42 | 5000 | 2 | 2500 |
| 72 | 4500 | 5 | 1000 |

Operational Parameters for In-Motion Mode Operation

| Maximum Fork Length (inches) | Capacity (lb) | e_{min} (lb) | n_{max} |
|------------------------------|---------------|----------------|-----------|
| 42 | 5000 | 5 | 1000 |

Standard Features:

- Auto Level Sensing and Correction Mechanism Contained in Junction Box
- DC Power Supply
- Hard-wired or Wireless Communication

Indicator:

- Rice Lake Weighing Systems indicating element Model 920i (NTEP CC No.01-088)
- or Model 420 Plus (NTEP CC No. 04-076)
- or Model VIRTUi (NTEP CC No. 04-058)
- or Model CLS680 (NTEP CC No. 19-021)
- or metrological equivalent NTEP certified instrument utilizing Rice Lake Weighing Systems CLS communication software

Load Cells Used:

- Two Revere Transducers Model 363-A5 (NTEP CC No. 87-063) or NTEP Certified Equivalent

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Marc Paquette
Chair, NCWM, Inc.

Gene Robertson
Chair, NTEP Committee
Issued: March 27, 2025

9011 South 83rd Street | Lincoln, Nebraska 68516

The National Council on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.



Rice Lake Weighing Systems

On-Board Weighing System / CLS

Application: The scale is designed for mounting onto the carriage of a lift truck (forklift). The scale is used to weigh commodities on pallets or skids for shipping when interfaced with any NTEP certified and compatible indicating element utilizing Rice Lake Weighing Systems CLS communication software.

Identification: A metal or adhesive label identification badge is attached to the side of the Weighing/Load Receiving Element.

Sealing: Set-up parameters and adjustments for the weighing element are made through the indicating element. The system incorporates an auto-level sensing angle correction mechanism inside the junction box. The junction box shall be sealed by threading a wire seal through two screws that fasten the cover. The indicator shall be sealed according to manufacturer specifications.

Operation: Depending on the configuration the instrument can either weigh statically or weigh in-motion. Switching between static weighing and in-motion weighing is not possible.

Static weighing: The forklift and vehicle must be at a standstill to capture weight values. Idling the vehicle engine during weighing operation is permissible. Instrumentation must be configured to prevent capture of weight while the load is in motion. The instrument will blank out when out-of-level exceeds the following conditions:

| Fork Length (Inches) | Capacity (lb) | e _{min} (lb) | Maximum Side to Side Tilt | Maximum Front Tilt | Maximum Rear Tilt |
|----------------------|---------------|-----------------------|---------------------------|--------------------|-------------------|
| 42 | 5000 | 2 | 3° | 7° | 3.5° |
| 42 | 5000 | 5 | 3° | 7° | 7° |
| 72 | 4500 | 5 | 3° | 7° | 7° |

Weighing in-motion: For in-motion weighing the instrument has a minimum weight defined. When the weight is below this minimum the instrument shows a live weight, giving the operator the opportunity to check zero. When exceeding the minimum weight value, the instrument will capture and hold the weight determined when motion, and level conditions are satisfied. The weight is held until the instrument is unloaded and the weight value drops below the minimum weight value. The instrument will blank out when out-of-level conditions exceed the following conditions:

| Fork Length (Inches) | Capacity (lb) | e _{min} (lb) | Maximum Side to Side Tilt | Maximum Front Tilt | Maximum Rear Tilt |
|----------------------|---------------|-----------------------|---------------------------|--------------------|-------------------|
| 42 | 5000 | 5 | 3° | 3° | 5° |

Testing: Depending on the operation it is recommended to test the instrument for field-testing according to the test procedures below.

Static weighing: For field testing, it is recommended that a standard-type skid be used as a load receiver. Test weights can then be loaded on the skid. Increasing/decreasing load and shift tests should be conducted. Out-of-level tests should be conducted using the conditions defined in the “Operation” paragraph above, to ensure that the device maintains accuracy when out-of-level, or the maximum possible out-of-level condition at that location, and that the leveling sensor is adjusted properly and inhibits the weighing operation when the system is positioned outside these limiting conditions. This may be accomplished by moving the lift truck to an out-of-level area and/or tilting the forks.

Weighing in-motion: For in-motion weighing it is recommended that a skid is used as a load receiver (if possible, a skid that is representative of a standard type skid). Static increasing load and shift tests should be conducted. Additionally, for testing the in-motion functionality it is recommended to test the instrument the same way as it is used, i.e. pick up the load, place the lift truck in motion and continue until the test load can be weighed and the weight value is displayed. Set down the load and check for the return to zero. Out-of-level tests should be conducted using the conditions defined in the “Operation” paragraph above, to ensure that the device maintains accuracy when out-of-level, or the maximum possible out-of-level condition at that location, and that the leveling sensor is adjusted properly and inhibits the weighing operation when the system is positioned outside these limiting conditions. This may be accomplished by moving the lift truck to an out-of-level area and/or tilting the forks.

Test Conditions: This certificate supersedes Certificate of Conformance 06-074A4 and is issued to decrease the verification scale interval from 5 lb to 2 lb when using the 42 inch forks in static mode only and to add 72 inch forks with a 5 lb scale verification interval for use in static mode only. The emphasis of the evaluation was on the design, operation and performance requirements.



Rice Lake Weighing Systems

On-Board Weighing System / CLS

The lift truck used for testing the 72 inch forks was a Class II, 36" x 16" carriage. Several static increasing/decreasing load and shift tests were conducted using test weights with the lift truck level and out-of-level 3° (5%) side to side and 7° front to back. A standard skid was used as the load receiver. A stationary frame was used to hold the 42 inch forks for the testing to reduce the verification scale interval. The permanence test for 42 inch forks with a verification scale interval of 2 lb included over 300 weighments in a 3-week period using pallets containing 1,000 lb, 3,000 lb, and 4,500 lb. Static increasing/decreasing load and shift tests were then repeated. No permanence testing was required to add the 72 inch forks. Previous test conditions are listed below for reference.

Certificate of Conformance 06-074A4: This certificate supersedes Certificate of Conformance 06-074A3 and is issued to document the relocation of the flexures for the Model CLS Onboard Weighing System. After a review of manufacturer information and according to Publication 14 technical policy, no further testing was deemed necessary. Previous test conditions are listed below for reference.

Certificate of Conformance 06-074A3: This certificate supersedes Certificate of Conformance 06-074A2 and is issued to evaluate the interaction of the Rice Lake Model CLS680 indicating element (NTEP CC 19-021) containing CLS communication software and the CLS on-board weighing system. Multiple increase/decrease testing was performed using a load cell simulator interfaced with the CLS680 Indicating element and the CLS "Auto Level Sensing and Correction Mechanism Junction Box". The indicating element will blank out when out-of-level conditions exceed 3° (5%) side to side, 5° (9%) backward and 3° (5%) forward this function was verified in all four directions. Previous test conditions are listed below for reference.

Certificate of Conformance 06-074A2: This certificate supersedes Certificate of Conformance 06-074A1 with the emphasis on the evaluation of the operation and performance requirements of in-motion weighing. The weighing/load receiving element was interfaced with the Rice Lake Weighing Systems, Model VIRTUi3 indicating element. Both static and in-motion field testing was performed on a 5000 x 5 lb lift truck. Increasing load, shift tests, and tilt tests were performed with the lift truck in both static and in-motion conditions. Previous test conditions are listed below for reference.

Certificate of Conformance 06-074A1: This certificate supersedes Certificate of Conformance 06-074 and was issued to increase the angle of tilt, front to back. Field-testing to extend the forward and back pitch from 3° to 7° was performed on a 5000 x 5 lb lift truck, passing multiple tests within acceptance tolerance. Out-of-level limiters would not allow indicator to produce readings past 7° as per specifications. No additional testing was deemed necessary.

Certificate of Conformance 06-074: The emphasis of the evaluation was on the design, operation and performance requirements. The scale was interfaced with Rice Lake Weighing Systems indicating element Model 920i (NTEP CC No. 01-088) utilizing 2 Revere transducers, Model 9363-A5-5K (NTEP CC No. 87-051A3). Separate tests were conducted utilizing Rice Lake Weighing Systems indicating Element Model 420 Plus (NTEP CC No. 04-076A2). Each indicating element was interfaced with the weighing element in a wireless and then wired configuration for testing purposes. The lift truck used for this test was a Class II, 36" x 16" carriage, 42" forks with 29" center-to-center fork spacing. Several static increasing/decreasing load and shift tests were conducted using test weights with the lift truck level and out-of-level 5 percent in all four directions. Tests were also conducted with the vehicle engine running. A standard skid was used as the load receiver. Test loads were distributed to simulate off center weight displacement. DC power supply tests were conducted. The permanence test for this device included over 300 weighments in a 3-week period using pallets containing 1,000 lb, 3,000 lb, and 4,500 lb. Static increasing/decreasing load tests were then repeated.

Evaluated By: D. Onwiler (NE) 06-074; M. Carlin (KS) 06-074A1; D. Flocken (NTEP) 06-074A2; J. Gibson (OH) 06-074A3; J. Gibson (NCWM) 06-074A4, (CN 11193); D. Flocken (NCWM) 06-074A5 (CN 11300)

Type Evaluation Criteria Used: NIST, Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices, 2025. NCWM, Publication 14: Weighing Devices, 2025.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: S. Patoray (NCWM), L. Bernetich (NCWM) 06-074; J. Truex (NCWM) 06-074A1, 06-074A2; D. Flocken (NCWM) 06-074A3, 06-074A4; J. Gibson (NCWM) 06-074A5



Rice Lake Weighing Systems
On-Board Weighing System / CLS

Example of Device:



Model CLS

