

# 1280 Enterprise™ Series

## 2020 Date Entry Workaround

---

**Product:** 1280 Enterprise Series indicator with firmware prior to version 2.00

**Issue:** Date cannot be set to a value after December 31, 2019 (123119)

**Solution:** Enter a 2019 date, then edit the date to 2020

---

If the date is invalid and a new date needs to be entered (i.e. when a battery is first installed or a unit has been defaulted), the firmware does not allow the date to be initially set to a value after December 31, 2019 (123119).

The workaround is to set the date to a valid 2019 date first (e.g. 042419), and then edit it to the current date (e.g. 042420).



**Note** *This date entry issue has only existed since the calendar year 2020 and has been addressed in firmware version 2.00.*

Reference the 1280 Enterprise Series Technical Manual (PN 167659) for complete product documentation.



Manuals can be viewed or downloaded from the Rice Lake Weighing Systems website at [www.ricelake.com/manuals](http://www.ricelake.com/manuals)



© Rice Lake Weighing Systems Specifications subject to change without notice.  
Rice Lake Weighing Systems is an ISO 9001 registered company.

230 W. Coleman St. • Rice Lake, WI 54868 • USA

U.S. 800-472-6703 • Canada/Mexico 800-321-6703 • International 715-234-9171 • Europe +31 (0)26 472 1319